

Clarius HD Scanner



Technical Specifications

Clarius HD Scanner Specifications

Model	Frequency	Max Depth	# Elements	Radius	Field of View	Pitch
C3 HD Convex	2–6 MHz	40 cm	192	45 mm	73°	300 µm
L7 HD Linear	4–13 MHz	11 cm	192	N/A	38 mm	200 µm
L15 HD Linear	5–15 MHz	7 cm	192	N/A	50 mm	260 µm
C7 HD Microconvex	3–10 MHz	15 cm	192	20 mm	112°	205 µm
EC7 HD Endocavity	3–10 MHz	15 cm	192	10 mm	164°	150 µm
PA HD Phased Array	1–4 MHz	40 cm	80	N/A	90°	250 µm

Imaging

Transmission

- 1 to 20 MHz waveforms
- Up to 20 continuous pulses
- Bi-polar output
- 10 to 80V peak-to-peak

Beamforming & Reception

- 8 Parallel Beamformers
- Synthetic Aperture Beamforming with virtual focal zones
- 60 MHz sampling rate @ 14 bits per channel

Post-processing

- Adaptive Speckle Reduction
- Edge enhancement
- Persistence

Total Input Dynamic Range

- 160dB

Automated Algorithms

- Time-Gain-Compensation (TGC)
- Frequency-Depth Adjustment
- Patient Contact Detection
- Needle Enhancement
- Motion Sensing

Imaging Modes

B-Mode	Yes
M-Mode	Yes
Power Doppler	Yes
Color Doppler	Yes
Pulsed-Wave Doppler	Optional
Needle Enhance (L7 HD/L15 HD)	Optional
Elastography	Optional
Harmonic Imaging	Cardiac

Internally Optimized Parameters

Clarius internally optimizes the following parameters to ensure the scanner is easy to use:

Frequency Range	1 to 16 MHz
Focal Zones Range	1 to 10
Compression Dynamic Range	30 to 90 dB
Reject	Yes
Sector Width Range	50% to 100%
Grey + Color Maps	Yes
Frame Rate	Up to 30 FPS

Interface Controls

- Depth
- Read Zoom
- 3 TGC sliders or Automated TGC
- Flip / Mirror
- Freeze
- Color / Power ROI
- Flow Speed
- Doppler Gate
- Doppler Correction Angle
- Doppler Steer
- Baseline
- Invert

Clinical Applications

- C3 HD**
- Abdominal
 - Bladder
 - Cardiac
 - Lung
 - MSK
 - OB/GYN
 - Prostate
 - Superficial

- L7 HD/
L15 HD**
- Breast*
 - Lung
 - MSK
 - Nerve
 - Ocular
 - Plastic Surgery
 - Small Parts (e.g. Thyroid)
 - Vascular

- C7 HD**
- Abdominal
 - Bladder
 - Cardiac
 - Lung
 - MSK
 - Small Parts
 - Speech Therapy

- EC7 HD**
- Early OB
 - IVF
 - Pelvic
 - Prostate

- PA HD**
- Abdominal
 - Bladder
 - Cardiac
 - Lung
 - OB/GYN
 - Superficial

Mobile Platforms

iOS iOS 11.0 or later

Android Android 4.4.2 (API 19) or greater and devices which have the following architectures: x64 and ARM. Devices must be compatible with Wi-Fi 802.11n and BLE.

Security and Encryption

Wi-Fi data channel TLS 1.2

Bluetooth AES128 and RSA4096

Data Management

Local Export JPG/PNG/DICOM/BMP

Cloud Export Yes

DICOM Store Optional

DICOM Worklist Optional

Connectivity

Wi-Fi 802.11n, dual band 2.4GHz & 5GHz

Bluetooth Bluetooth Low Energy 4.1

Measurements and Calculations

TOOLS

Distance Yes

Trace Yes

Ellipse Yes

Heart Rate Yes

Time Yes

Velocity Yes

Volumes Yes → Manual/Automated

CALCULATION PACKAGES

Obstetrics BDP, HC, FL, AC, CRL, GS, AFI. EFW from GA. Hadlock tables, HR

IVF Left/Right Follicles

Vascular Stenosis, ICA, ECA, CCA

Cardiac Basic IVS, Basic LVID, LV/RV Tracing, HR, Basic LVPW, EPSS

* Available via Hologic's Viera Breast Performance Package

MEASUREMENT ACCURACY

Lateral Distance

Relative Error < +- 2%

Minimum Range <= 0.2 mm

Maximum Range >= 24 cm

Axial Distance

Relative Error < +- 2%

Minimum Range <= 0.2 mm

Maximum Range >= 24 cm

Doppler Sensitivity

C3 HD/PA HD Depth Sensitivity: 10.0 cm
Flow Sensitivity:
0.1 mL/sec at a depth of 5.0 cm

L7 HD/L15 HD Depth Sensitivity: 5.0 cm
Flow Sensitivity:
0.1 mL/sec at a depth of 5.0 cm

C7 HD Depth Sensitivity: 7.0 cm
Flow Sensitivity:
0.1 mL/sec at a depth of 5.0 cm

EC7 HD Depth Sensitivity: 6.0 cm Flow
Sensitivity:
0.1 mL/sec at a depth of 5.0 cm

Mechanical

- Enclosure**
- Light weight magnesium
 - Durable
 - IP67 rated for probe and battery separately, rated for 1 metre immersion for 30 minutes

SCANNER DIMENSIONS AND WEIGHT (WITH BATTERY)

C3 HD Dimensions: 164 x 78 x 38 mm
Weight: 392 g

L7 HD Dimensions: 165 x 78 x 38 mm
Weight: 364 g

L15 HD Dimensions: 165 x 78 x 38 mm
Weight: 372 g

C7 HD Dimensions: 169 x 78 x 38 mm
Weight: 366 g

EC7 HD Dimensions: 328 x 78 x 38 mm
Weight: 410 g

PA HD Dimensions: 166 x 78 x 38 mm
Weight: 368 g

ACCESSORIES DIMENSIONS AND WEIGHT

Battery Dimensions: 27 x 72 x 32 mm
Weight: 68 g

Battery charger (without plug adapter) Dimensions: 48 x 74 x 49 mm
Weight: 60 g

Battery, Charging, and Bootup

Battery Life ~60 min scanning

Standby ~7 days idle

Charge Time ~90 min

Max Scan Time Per Exam ~30 min

Bootup Platform dependent, generally less than 30 seconds

Cleaning

Tested without adverse effects

- Accel® PREvention™ Wipes
- Accel® TB Wipes
- CaviWipes
- CIDEX® OPA
- MetriCide™ OPA Plus High Level Disinfectant Solution
- Sani-Cloth® HB Germicidal Disposable Wipe
- Sani-Cloth® Plus Germicidal Disposable Cloth
- Tristel Trio Wipes System
- Virox™AHP® 5 RTU Wipes

Standard Configuration

- Scanner
- 1 Rechargeable Battery
- 1 Charger with global AC Adapter
- Access to Clarius Cloud platform

WARRANTY

Included 3 years limited warranty

- Optional** Clarius Care - 1/2/3 years
- Accidental damage
 - Uptime
 - RMA shipping
 - Hospital theft

Standards Compliance

IEC 60601-1:2012, Medical Electrical Equipment – Part 1: General requirements for basic safety and essential performance

IEC 60601-1-2:2014, Medical Electrical Equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests

IEC 60601-2-37:2007, Medical Electrical Equipment – Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment

NEMA UD-2, Acoustic Output Measurement Standard For Diagnostic Ultrasound Equipment, Revision 3

NEMA UD-3, Standard for Real-Time Display of Thermal and Mechanical Acoustic Output Indices on Diagnostic Ultrasound Equipment, Revision 2

IEC 60601-1-12:2014, Medical electrical equipment – Part 1-12: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment

FCC 47CFR Part 15, Radio frequency devices

ETSI EN 300 328:V2.1.1:2016-11 – Electromagnetic compatibility and Radio spectrum Matters (ERM)

ETSI EN 301 489-1:V2.2.0:2017-03 – Electromagnetic compatibility and Radio spectrum Matters (ERM)

ETSI EN 301 489-17:V3.1.1:2017-02 – Electromagnetic compatibility and Radio spectrum Matters (ERM)

ISO 10993-1:2009, Biological evaluation of medical devices

IEC 60529:2013, Degrees of protection provided by enclosures (IP Code)

IEC 62133:2012, Secondary cells and batteries containing alkaline or other non-acid electrolytes – Safety requirements for portable sealed secondary cells, and for batteries made from them, for use in portable applications

UN 38.3, Transport of dangerous goods – Classification procedures, test methods and criteria relating to class 9 – Lithium metal and lithium ion batteries

About Us

Clarius Mobile Health was founded by experienced innovators who have played an instrumental role in the ultrasound industry. Our developers were the brains behind the first PC-based platform for ultrasound research. They also introduced the first touch screen ultrasound system with a simplified user interface.

We started with a simple mission: to enable more clinicians to use ultrasound to improve patient care. Thanks to the power of smart phones, advanced technology and decades of collective ultrasound experience, we've developed a high quality, Point-and-Shoot Ultrasound™ scanner that works with your smart device.

Clarius Mobile Health
350-3605 Gilmore Way
Burnaby, BC, V5G 4X5,
Canada
hello@clarius.com
Phone: 1-778-800-9975